

**COMPARATIVE ANALYSIS OF DOSIMETRY READINGS
SCATTERBAN RADIATION BARRIER VS NO RADIATION BARRIER**

MONTH 1999	# OF FLUORO MINUTES	#OF FLUORO PROCEDURES	MREM DOSIMETRY FINDINGS PRIOR TO IMPLEMENTATION OF SCATTERBAN SYSTEM		
			DEEP	EYE	SHALLOW
January	140.02	83	510	510	510
February	84.39	75	240	240	240
March	99.33	70	470	470	470

MONTH 1999	# OF FLUORO MINUTES	#OF FLUORO PROCEDURES	MREM DOSIMETRY FINDINGS AFTER IMPLEMENTATION OF PARTIAL SCATTERBAN SYSTEM		
			DEEP	EYE	SHALLOW
May	136.08	118	166	166	161
June	125.45	123	24	24	23
July	107.39	115	37	37	37

MONTH 1999	# OF FLUORO MINUTES	#OF FLUORO PROCEDURES	MREM DOSIMETRY FINDINGS AFTER IMPLEMENTATION OF COMPLETE SCATTERBAN SYSTEM		
			DEEP	EYE	SHALLOW
August	153.07	123	2	2	1
September	141.59	131	3	4	4
October	149.46	136	2	2	1
November	199.36	129	2	2	2
December	114.02	125	1	1	M

MREM - Minimal Radiation Equivalent Measure
M - Dose equivalent below the minimal measurable quality